



**Durable,³
Dependable,
and Accurate^{1,4}**

The New Face of Genius™ 3 Tympanic Thermometer Industry Standard Tested Performance, Improved Durability^{1,3}

- Peak Select technology will display accurate temperature in 1 to 2 seconds.¹
- Durable design reduces the need for recalibration.^{1,2}
- Disposable single-use probe covers help prevent cross-contamination.¹



PRODUCT SPECIFICATIONS

Accuracy

- +/- 0.3°C⁴

Measurement Response Time

- 1–2 seconds

Ambient Operating Range

- 16°C–33°C (60.8°F–91.4°F)

Temperature Measurement Range

- 33°C–42°C (91.4°F–107.6°F)

Storage Temperature Range

- -25°C–55°C

Battery Life

- 15,000 temperatures (3 AAA batteries)

Anti-Theft System

- Wall mount bracket/Genius™ mobile cart

Compliant with Standards

- ISO 80601-2-56:2017⁴
- IEC 60601-1-2:2005/AMD1:2012³
- IEC 60601-1-2:2014⁵
- IEC 60601-1-11:2015 Ed.2³

Degree of Protection Against Ingress of Fluids

- DripProof - IP22

Warranty

- 3 year manufacturer warranty

Genius™ 3 Thermometer: Robust, dependable and accurate¹

- **Redesigned plastic housing allows for an IP22 rating for protection against dust and liquid ingress.⁶**
- Peak Select technology takes multiple readings and selects the highest temperature for display in 1 to 2 seconds.¹
- Durable design reduces the need for recalibration.^{1,2}
- Disposable single-use probe covers help prevent cross-contamination.

Order Information

Code	Description	Ship Case
Z303013	Genius™ 3 Thermometer with Base	1
Z303030	Genius™ Probe Cover	2112
Z303058	Genius™ Locking Mount Wall	1
Z303097	Genius™ Checker/Calibrator	1

1. Based on internal report #P-4359-049, revision 3, Ear Thermometer Simulated Use Testing for Accuracy. March 2007.

2. Based on internal report #P-4359-049, revision 2, Ear Thermometer Simulated One Year Use Testing for Accuracy. June 2006.

3. Based on internal test report #IEC 60601-1, Medical electrical equipment — Part 1: General requirements for basic safety and essential performance. February 2018.

4. Based on internal test report #ISO 80601-2-56, Medical electrical equipment — Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement. February 2018.

5. Based on internal test report #COVI0147, revision 1, Medtronic Genius 3. July 2017.

6. Based on internal test report #IEC 60601-1-11, Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance. Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment. February 2018.